

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 12 April 2018 Instrument Serial Number: 100351

Instrument Location: Fairbanks Airport PD

Supervisor Name and Agency: Vance E Enderle - Fairbanks Airport PD

Describe the Status Change or Issue with the Instrument:

New instrument put in service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 12 April 2018

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740