

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/17/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.72: 0.077  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 1183 psi

BLANK TEST	0.000	15:42
INTERNAL STANDARD	VERIFIED	15:42
EXTERNAL STANDARD	0.078	15:43
BLANK TEST	0.000	15:44
EXTERNAL STANDARD	0.076	15:44
BLANK TEST	0.000	15:45
EXTERNAL STANDARD	0.075	15:45
BLANK TEST	0.000	15:46
EXTERNAL STANDARD	0.081	15:46
BLANK TEST	0.000	15:47
EXTERNAL STANDARD	0.076	15:47
BLANK TEST	0.000	15:48

Average = 0.0772  
Std Dev = 0.0024

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C	PASSED
Breath Tube = 44.6°C	PASSED

### PUMP INFO

Flow Rate = 4.657 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only: Troubleshooting occurs prior to  
new passing voc. BMB 12/27/19