

VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument
State of Alaska*

Serial #: 100690

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:06/14/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.73: 0.079
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 1221 psi

BLANK TEST	0.000	00:38
INTERNAL STANDARD	VERIFIED	00:39
EXTERNAL STANDARD	0.079	00:39
BLANK TEST	0.000	00:40
EXTERNAL STANDARD	0.077	00:40
BLANK TEST	0.000	00:41
EXTERNAL STANDARD	0.079	00:41
BLANK TEST	0.000	00:42
EXTERNAL STANDARD	0.080	00:42
BLANK TEST	0.000	00:43
EXTERNAL STANDARD	0.079	00:43

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 4.461 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------