

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100374

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/10/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.51: 0.079
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 147 psi

BLANK TEST	0.000	20:45
INTERNAL STANDARD	VERIFIED	20:45
EXTERNAL STANDARD	0.077	20:45
BLANK TEST	0.000	20:46
EXTERNAL STANDARD	0.078	20:46
BLANK TEST	0.000	20:47
EXTERNAL STANDARD	0.077	20:47
BLANK TEST	0.000	20:48
EXTERNAL STANDARD	0.077	20:48
BLANK TEST	0.000	20:49
EXTERNAL STANDARD	0.077	20:49
BLANK TEST	0.000	20:50

Average = 0.0772
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 42.0°C PASSED

PUMP INFO

Flow Rate = 4.300 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only: Instrument being removed from
service. BMB 10/23/19