

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100682

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 04/27/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.66: 0.079

LOT #: AG735001

EXPIRATION: 12/16/2019

TANK PRESSURE: 1182 psi

BLANK TEST                    0.000   17:11  
INTERNAL STANDARD            ERROR   17:12

Internal Standard Error

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C            PASSED  
Breath Tube       = 47.5°C           PASSED

### PUMP INFO

Flow Rate = 4.658 L/M               PASSED

### DETECTOR INFO

PUMP ON                               PASSED

PUMP OFF                              PASSED

### FILTER INFO

Filter 1                                PASSED

Filter 2                                PASSED

Filter 3                                PASSED

INTERNAL STANDARD                   PASSED