

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100380

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/17/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.91: 0.080

LOT #: AG310901

EXPIRATION: 04/19/2025

TANK PRESSURE: 429 psi

BLANK TEST 0.000 10:10

INTERNAL STANDARD VERIFIED 10:10

EXTERNAL STANDARD 0.085 10:10

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C

PASSED

Breath Tube = 43.6°C

PASSED

PUMP INFO

Flow Rate = 4.542 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

PASSED

File only
CMB-3/6/25