

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 7 Aug 2017 Instrument Serial Number: 100358
Instrument Location: Whittier Police Dept Gardwood
Supervisor Name and Agency: Dave Schubeil Whittier P.D.

Describe the Status Change or Issue with the Instrument:

out of service to test modem at crime lab

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 7 Aug 2017 / 1657 hrs

Out of Service Date/Time: 7 Aug 2017

For Use by SCDL

Additional Notes

Removed from service in Gardwood + transported to SCDL to test the functionality of the modem. Instrument uploaded @ SCDL without issue. Returned to service in Gardwood and will troubleshoot on site modem here. COB 8/8/17