

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100413Date: 6/24/11

1. Wet Bath Simulator Solution

- a. Level 0.100
 b. Manufacturer Repro
 c. Lot 10003
 d. Expiration Date 9/8/12

2. Wet Bath Simulator

- a. Serial Number DR4161
 b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst

C. O'Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dnt: 100413

Date: 06/24/2011
Time: 09:52:35

SUPERVISOR NAME:
COLLEEN S OBRYANT
CERTIFICATE #: 6237

LOT #: REPC0100
BOTTLE #: 10003
EXPIRATION: 09/08/2012

Ca	=	0.1000		
CAL	=	0.979992	0.800	<= CAL < 1.200
b1	=	0.0002	0.0000	<= b1 < 0.0040
b2	=	0.0041	0.0010	<= b2 < 0.0100
b3	=	0.0002	0.0000	<= b3 < 0.0040
Xq	=	0.0917	0.0500	<= Xq < 0.2000
a21	=	1.130309	1.050	<= a21 < 1.300
a31	=	0.436579	0.300	<= a31 < 0.800

COB-2

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100413\B-100413_ECF-20140

2/18/2014