

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 9/11/21 Instrument Serial Number: 100675

Instrument Location: Bethel PD - Patriot

Supervisor Name and Agency: Justin Ulak Bethel PD

Describe the Status Change or Issue with the Instrument:

Failed VOC's Detector Overflow status.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 9/11/21 2107 hrs.

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740