

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100384

Date: 11/03/2023
Time: 15:05:48

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
DEREK WALTON
OPERATOR'S #: 9508

DEPT/AGENCY:
NON-DRINKING SUBJECT
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.34: 0.078
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK	PASSED	15:08
BLANK TEST	0.000	15:08
INTERNAL STANDARD	VERIFIED	15:08
EXTERNAL STANDARD	0.078	15:08
BLANK TEST	0.000	15:09
SUBJECT SAMPLE (Vol=2.33L)	0.000	15:10
BLANK TEST	0.000	15:11
EXTERNAL STANDARD	0.078	15:11
BLANK TEST	0.000	15:12

DM

