

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100413

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/12/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.81: 0.077
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 438 psi

| | | |
|-------------------|----------|-------|
| BLANK TEST | 0.000 | 13:28 |
| INTERNAL STANDARD | VERIFIED | 13:28 |
| EXTERNAL STANDARD | 0.075 | 13:29 |
| BLANK TEST | 0.000 | 13:30 |
| EXTERNAL STANDARD | 0.075 | 13:30 |
| BLANK TEST | 0.000 | 13:31 |
| EXTERNAL STANDARD | 0.075 | 13:31 |
| BLANK TEST | 0.000 | 13:32 |
| EXTERNAL STANDARD | 0.076 | 13:32 |
| BLANK TEST | 0.000 | 13:33 |
| EXTERNAL STANDARD | 0.076 | 13:33 |
| BLANK TEST | 0.000 | 13:34 |

Average = 0.0754
Std Dev = 0.0005

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

| | |
|-------------------------|--------|
| Sample Chamber = 48.7°C | PASSED |
| Breath Tube = 48.1°C | PASSED |

PUMP INFO

| | |
|-----------------------|--------|
| Flow Rate = 4.862 L/M | PASSED |
|-----------------------|--------|

DETECTOR INFO

| | |
|----------|--------|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |

FILTER INFO

| | |
|----------|--------|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

| | |
|-------------------|--------|
| INTERNAL STANDARD | PASSED |
|-------------------|--------|

File Only Instrument Removed
from Service 1/11/21 BMB 2/17/2021
See 2/17/2021 Change in Instrument
Status form for more information.