

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100414

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date:01/06/2022

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.72: 0.079

LOT #: AG113402

EXPIRATION: 05/14/2023

TANK PRESSURE: 217 psi

BLANK TEST 0.000 14:13

INTERNAL STANDARD VERIFIED 14:13

EXTERNAL STANDARD INTERFERENCE 14:13

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C PASSED

Breath Tube = 46.0°C PASSED

### PUMP INFO

Flow Rate = 4.859 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED