

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 8/17/16 Instrument Serial Number: 100674

Instrument Location: Tuneau PD

Supervisor Name and Agency: Colleen O'Bryant, SCDL

Describe the Status Change or Issue with the Instrument:

Sterling Salisbury (JPD) called to say this instrument had given two standard out of range status messages on 8/16/16 following subject tests as well as two blank errors on 8/17/16 following attempted VOCs. I told Sterling to purge the sample chamber and the room the DataMaster was in, the messages could be due to environmental factors. Sterling purged the instrument and the room and was no longer having the issue. The instrument is still in service. COB 8/17/16

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740