

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: ^{6/19/23} Click or tap to enter a date.

Instrument #: ¹⁰⁰⁷⁰⁰ Click or tap here to enter text.

Instrument Location: Click or tap here to enter text. ^{Palmer PD}

Supervisor Name and Agency: Click or tap here to enter text. ^{Sgt Luke Szipszky}

Describe the Status Change or Issue with the Instrument:

Screen would freeze at "signature box" and not allow the test to complete. Happened several times over weekend.
worked right before disconnecting so I ran a VOC.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

^{OUT}

~~In~~ Service Date/Time: Click or tap here to enter text. ^{6/19/23 1800}

~~Out of Service~~ Date/Time: Click or tap here to enter text.

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor