

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 7 May 2015 Instrument Serial Number: 100342

Instrument Location: JBEA, AK RID 5517

Supervisor Name and Agency: Scott Gardner, 6222 JBEA

Describe the Status Change or Issue with the Instrument:

placed into service as a temporary replacement.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 7 May 2015 / 1358

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740