

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100689

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 10/15/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.32: 0.081  
LOT #: AG923401  
EXPIRATION: 08/22/2021  
TANK PRESSURE: 1226 psi

BLANK TEST	0.000	00:01
INTERNAL STANDARD	VERIFIED	00:01
EXTERNAL STANDARD	0.079	00:01
BLANK TEST	0.000	00:02
EXTERNAL STANDARD	INTERFERENCE	00:02

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C	PASSED
Breath Tube = 44.6°C	PASSED

### PUMP INFO

Flow Rate = 4.299 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED