

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 9/27/2024

Instrument #: 100365

Instrument Location: Skagway Police Department

Supervisor Name and Agency: James Michels, Skagway Police

Describe the Status Change or Issue with the Instrument:

Out of Service – Internal Standard Error, replaced with DataMaster 100368.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 9/27/2024 9:15 am

For Use by SCDL

Additional Notes

Click or tap here to enter text.

See e-mail below for more information. DataMaster will be dropped off in person on 10/08/2024.
MNC 09/27/2024

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/27/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.66: 0.079
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 714 psi

BLANK TEST 0.000 09:10
INTERNAL STANDARD ERROR 09:10

Internal Standard Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 44.4°C PASSED

PUMP INFO

Flow Rate = 4.578 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

Carlock, Megan (DPS)

From: James Michels <J.Michels@skagway.org>
Sent: Monday, September 23, 2024 11:38 AM
To: Carlock, Megan (DPS)
Subject: RE: [External Email] RE: Skagway DataMaster

Follow Up Flag: Follow up
Flag Status: Completed

Sounds good, hopefully it will be here before I leave for the Re-Cert on October 4th, if so I will bring the old one back to you in person.....Thanks, James



James Michels
Skagway Police Dept.
PO Box 518
Skagway, AK 99840
j.michels@skagway.org
907-612-0255

From: Carlock, Megan (DPS) <Megan.Carlock@alaska.gov>
Sent: Monday, September 23, 2024 10:26 AM
To: James Michels <J.Michels@skagway.org>
Subject: [External Email] RE: Skagway DataMaster

[ATTENTION: This Email was received from outside the Municipality]

Good morning,

I took a look at your DataMaster records, and I will send you a new instrument today. It looks like the internal standard reading is persistently high which is causing these errors from time to time.

I will send you the tracking information for the new instrument once it is shipped. Please let me know if you have any questions.

Thank you,

Megan Carlock (Crouse)



Forensic Scientist II – Chemistry Section
State of Alaska Scientific Crime Detection Laboratory
907-269-3005

From: James Michels <J.Michels@skagway.org>
Sent: Wednesday, September 18, 2024 3:04 PM
To: DPS SCDL Toxicology (DPS sponsored) <dps.scdl.toxicology@alaska.gov>
Subject: Skagway DataMaster

CAUTION: This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Good Afternoon,

I came to work an hour ago and was told DataMaster 100365 was being grumpy, multiple status messages. I ran a test and it worked fine. I attached the status message sheets to this email. Any thoughts? Thanks....James



James Michels
Skagway Police Dept.
PO Box 518
Skagway, AK 99840
j.michels@skagway.org
907-612-0255

[unsubscribe](#)

[unsubscribe](#)