

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form – DMCAF2011R1		
	Issuing Authority – Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100366

Date: 3/22/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Guth
- c. Lot 9380
- d. Expiration Date 12/17/11

2. Wet Bath Simulator

- a. Serial Number DR4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100366

Date: 03/22/2011
Time: 11:48:02

SUPERVISOR NAME:
COLLEEN S OBRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 9380
EXPIRATION: 12/17/2011

Ca	=	0.1000		
CAL	=	0.985254	0.800	<= CAL < 1.200
b1	=	0.0000	0.0000	<= b1 < 0.0040
b2	=	0.0054	0.0010	<= b2 < 0.0100
b3	=	0.0000	0.0000	<= b3 < 0.0040
Xq	=	0.1000	0.0500	<= Xq < 0.2000
a21	=	1.200050	1.050	<= a21 < 1.300
a31	=	0.503220	0.300	<= a31 < 0.800