

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 02/01/2018 Instrument Serial Number: 100415

Instrument Location: Sitka PD

Supervisor Name and Agency: Charles Foster - SCDL

Describe the Status Change or Issue with the Instrument:

---

---

---

---

---

---

---

---

---

---

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

-----  
**For Use by SCDL**

**Additional Notes**

On the afternoon of 01/31/2018 it was discovered that the tank used in 100415 from 06/01/2017 (in service) to current date was AG608504 rather than AG634001 as indicated by the VOCs and subject tests from 06/01/2017 to 01/03/2018. This was discovered by having Detective Ferguson (BTO 9161) and Sgt. Johnson (BTO 4639) remove tank AG608504 from the instrument and see that it didn't match the serial number listed on the instrument screen (AG634001). Sgt. Johnson initiated a tank change and VOC with the correct external standard lot number on 01/31/2018.

Both tank lot numbers and expiration dates are current and valid for the dates the instrument was used.

AG608504, exp 3/25/2018 was used from 06/01/2017 to 01/31/2018 but was listed as AG634001, exp 12/15/2018 on VOCs and subject tests.

- Charles Foster 2/1/2018

*CWF*

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740