

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100682

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/09/2025

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C

PASSED

Breath Tube = 47.3°C

PASSED

PUMP INFO

Flow Rate = 4.926 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

FAILED

Diagnostic Check Failed

File Only jmm 06/18/25