

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100677

Date: 6/13/16

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Instrument returned for various messages (undefined error and unable to load ticket). At Scott instrument gave various software messages and instrument would not upload.

2. Visual Inspection

3. Troubleshooting/Repair

Attempted to install a new copy of the ticket db which may have failed. Consulted Scott (NPTAS) for assistance. Unable to load a new ticket db via windows explorer/disk on a chip. Instrument frozen at NPTAS black screen.

4. VOC (attached) NA

5. Internal Standard value $\leq 3.00\%$ Yes No NA

6. Verify software version NA

Return to Factory for Repair

Ready For Calibration

Ready For Certification

