

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 8/17/2022

Instrument #: 100384

Instrument Location: Sand Point PD

Supervisor Name and Agency: Colleen O'Bryant

Describe the Status Change or Issue with the Instrument:

100384 was placed in service at Sand Point PD by Richard Lowery (100384 2021.11.19 Change In Status). At the time of in service, a tank change was not recorded, therefore the lot number and expiration date of the tank installed in the instrument did not match the lot number and expiration date stored in the instrument from a time period of 2021.11.19 – 2022.07.08. All tests completed during that time frame have the incorrect external dry gas tank information of "AG018102 Exp 06/29/2022" on the records. SCDL records show external dry gas standard AG1130402 Exp 5/14/2023 was the only tank mailed to Sand Point PD in July 2021. VOC records will be corrected for lot number and expiration date. See attached emails for additional documentation.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: N/A

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

100384 2021.11.19 - 2022.07.08 Change In Status

COB 2022.08.17 Page 1 of 8

O'Bryant, Colleen S (DPS)

From: Barnett, Brandi M (DPS)
Sent: Wednesday, August 17, 2022 1:42 PM
To: O'Bryant, Colleen S (DPS)
Subject: FW: Need a new external tank

Brandi M. Barnett
Quality Assurance Manager
Alaska Scientific Crime Detection Laboratory
907-269-5695

From: Barnett, Brandi M (DPS)
Sent: Monday, July 11, 2022 6:59 AM
To: O'Bryant, Colleen S (DPS) <colleen.obryant@alaska.gov>
Subject: FW: Need a new external tank

Reading back through this I realized that Sand Point likely put a tank into their instrument before turning it on and had an incorrect lot and expiration date since it went into service. The inservice date is 11/19/2021 at 11:20. There is no tank change on that day. Can you please look into this when you get back and fill out a change in instrument status. They also made a mess of their testing and we can discuss this tomorrow as well 😊.

Brandi M. Barnett
Quality Assurance Manager, Chemistry Supervisor
Scientific Director - Breath and Blood Alcohol Program
Alaska Scientific Crime Detection Laboratory
907-269-5695

From: Lee, Tatiana A (DPS)
Sent: Friday, July 8, 2022 10:55 AM
To: Barnett, Brandi M (DPS) <brandi.barnett@alaska.gov>
Subject: RE: Need a new external tank

Hey Brandi,

Richard Lowery held a BTO class and had this students take the paper exam prior getting approval. He doesn't think they have enough internet power to support online testing. Would you like them to wait until they regain their internet or okay the paper exams?



Tatiana Lee

TruhqvIF#WhfkqIFbq#, #F khp lwu| #HyIghqfh#
Døvnd#7 f l hq w i f #F u p h#G h w f w r q#D e r u d r u l #
7 ; 3 8 #G u #P d w i q#D x w k h u #N l j #M u #D y h q x h #D q f k r u d j h #D ø v n d # < 8 3 : #
#G l h f w # < 3 : , # 9 < 0 8 3 < 9 #

From: Richard Lowery <rlowery@sandpointak.org>
Sent: Friday, July 8, 2022 10:53 AM

To: Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov>

Subject: Re: Need a new external tank

It's in and out right now, I do not believe we currently have enough bandwidth to support online testing. Our main internet is out and we are using an old telephone modem for emails

Richard Lowery

Task Force Officer

Sand Point Police Department

P.O. Box 423 Sand Point AK, 99661

P. (907) 383-3700

F. (907) 383-5496

This e-mail message may contain confidential or legally privileged information and is intended only for the use of the intended recipient(s). If you are not the intended recipient, you are hereby notified that any unauthorized review, disclosure, dissemination, distribution, duplication or taking of any action in reliance on the information herein is strictly prohibited. If you are not the intended recipient, please contact the sender by reply email and destroy all copies of the original message.

On Fri, Jul 8, 2022 at 10:48 AM Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov> wrote:

All exams must be taken through the portal. Is your internet currently down?



Tatiana Lee

Iruhqvlf#whfkqfllqg# #Fkhp lwu| ##lyghqfh

D@vnd#Vf lhgwlif#F up h#G hwhfwlrq#Dderudru|

7;38#Gu#P duwq#Dxwkhun lqj #Mu#Dyhqxn#Dqfkrudjh#D@vnd#<83 : #

#G lhfvr#<3 : #59<083<9

From: Richard Lowery <rlowery@sandpointak.org>

Sent: Friday, July 8, 2022 10:46 AM

To: Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov>

Subject: Re: Need a new external tank

Our internet is intermittent at this time, I had them do the paper exam, will that be okay?

Richard Lowery

Task Force Officer

Sand Point Police Department

P.O. Box 423 Sand Point AK, 99661

P. (907) 383-3700

F. (907) 383-5496

This e-mail message may contain confidential or legally privileged information and is intended only for the use of the intended recipient(s). If you are not the intended recipient, you are hereby notified that any unauthorized review, disclosure, dissemination, distribution, duplication or taking of any action in reliance on the information herein is strictly prohibited. If you are not the intended recipient, please contact the sender by reply email and destroy all copies of the original message.

On Fri, Jul 8, 2022 at 10:28 AM Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov> wrote:

Please complete the following form so I can create the course in ACADIS.



Tatiana Lee

Iruhqvlf#WhfkqIfldq#, #F khp lwu| #HyLghqfh

Døvnd#VfInqwilf#Fup h#Ghwhfwlrq#Døvndwru|

7;38#GuP dwtq#Dxwku#Nlj#Mu#Dyhqxh#Dqfkrudjh/#Døvnd#<83:#

#3lhfw#<3:#59<083<9

From: Richard Lowery <rlowery@sandpointak.org>
Sent: Friday, July 8, 2022 10:17 AM
To: Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov>
Subject: Re: Need a new external tank

Where is the course eval for the students who have just completed training?

Richard Lowery

Task Force Officer

Sand Point Police Department

P.O. Box 423 Sand Point AK, 99661

P. (907) 383-3700

F. (907) 383-5496

This e-mail message may contain confidential or legally privileged information and is intended only for the use of the intended recipient(s). If you are not the intended recipient, you are hereby notified that any unauthorized review, disclosure, dissemination, distribution, duplication or taking of any action in reliance on the information herein is strictly prohibited. If you are not the intended recipient, please contact the sender by reply email and destroy all copies of the original message.

On Fri, Jul 8, 2022 at 10:11 AM Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov> wrote:

Awesome, thanks! Let me know if you need anything else.



Tatiana Lee

Irnhqvlf#Whfkqfllq#,#khp lwl|#Hyghqfh

D@vnd#7f1hqwllf#F up h#Ghwhfwlrq#Dderudwru|

7;38#G u#P duq#Dxwku#Nlj#M#Dyhxh#Dqfkrudjh/#D@vnd#<83:#

#G lhf#<3:,#59<083<9

From: Richard Lowery <rlowery@sandpointak.org>
Sent: Friday, July 8, 2022 10:07 AM
To: Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov>
Subject: Re: Need a new external tank

Tatiana,

My mistake, I suppose I did not enter in the appropriate information when I changed the machines out last year. I reentered the correct information and we are back on track.

Thanks,

Richard Lowery

Task Force Officer

Sand Point Police Department

P.O. Box 423 Sand Point AK, 99661

P. (907) 383-3700

F. (907) 383-5496

This e-mail message may contain confidential or legally privileged information and is intended only for the use of the intended recipient(s). If you are not the intended recipient, you are hereby notified that any unauthorized review, disclosure, dissemination, distribution, duplication or taking of any action in reliance on the information herein is strictly prohibited. If you are not the intended recipient, please contact the sender by reply email and destroy all copies of the original message.

On Fri, Jul 8, 2022 at 8:10 AM Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov> wrote:

Officer Lowery,

Our records shows that we sent Sand Point PD a tank last July 2021. The lot number for this tank is AG113402 and this lot expires on 05-14-2023. Could you verify whether you received this tank?

Officer Nierman was notified of this shipment last year, I understand that he's no longer with your agency so please let me know ASAP so I can get one sent out today.



Tatiana Lee

Irnhqvlf#nhfkqfldq#, #Fkhp lwa|##lylghqfh

D@vnd#vflhqvllf#Fup h#Shwhfwlrq#Dderudru|

7;38#Su#P dwtq#Dxwku#Nlj#Mu#Dyhqxh#Dqfkrudjh,#D@vnd#<83 :#

#G lhf#<3 : ,#59<083<9

From: Richard Lowery <rlowery@sandpointak.org>
Sent: Thursday, July 7, 2022 5:29 PM
To: Lee, Tatiana A (DPS) <tatiana.lee@alaska.gov>
Subject: Need a new external tank

CAUTION: This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Tatiana,

We are in need of a new dry gas tank. Ours expired on 06/29/2022. I currently have two new officers in BTO training as of today and did not realize our tank had expired. Could I please get a new tank sent off to me? I know this is super short notice but I will only be on the island until Wednesday of next week before going out on medical leave. Also, I will send the new operator form when these two have completed their training.

Thanks in advance,

Richard Lowery

Task Force Officer

Sand Point Police Department

P.O. Box 423 Sand Point AK, 99661

P. (907) 383-3700

F. (907) 383-5496

This e-mail message may contain confidential or legally privileged information and is intended only for the use of the intended recipient(s). If you are not the intended recipient, you are hereby notified that any unauthorized review, disclosure, dissemination, distribution, duplication or taking of any action in reliance on the information herein is strictly prohibited. If you are not the intended recipient, please contact the sender by reply email and destroy all copies of the original message.