

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 1/9/15 Instrument Serial Number: 100404

Instrument Location: UAF PD

Supervisor Name and Agency: Colleen O'Bryant / SCDL

Describe the Status Change or Issue with the Instrument:

See attached email for description of instrument issue.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

I spoke with Sam Beaudreault of UAF PD about the fatal system error that occurred. We agreed to leave the instrument in service and monitor the issue.
COB 1/12/15

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

O`Bryant, Colleen S (DPS)

From: DPS SCDL Toxicology (DPS sponsored)
Sent: Thursday, January 08, 2015 10:24 AM
To: Bolz, Nita J (DPS); O`Bryant, Colleen S (DPS); Foster, Charles R (DPS); Barnett, Brandi M (DPS)
Subject: FW: UAFPD Datamaster

From: Samuel Beaudreault [sabeaudreault@alaska.edu]
Sent: Wednesday, January 07, 2015 7:51 PM
To: DPS SCDL Toxicology (DPS sponsored)
Subject: Fwd: UAFPD Datamaster

Here is some more information from the officer that had the problem, about Datamaster dmt 100404 at the UAF Police Department,

It looks like it might be an actual issue.

The machine is turned off right now due to construction near it.
Let me know what you suggest.

Thank you!

----- Forwarded message -----

From: Brown, Travis A (DOT) <travis.brown@alaska.gov>
Date: Wed, Jan 7, 2015 at 7:35 PM
Subject: Re: UAFPD Datamaster
To: Samuel Beaudreault <sabeaudreault@alaska.edu>

Ofc. Beaudreault,

Ofc. Hall had same error weeks prior. He didn't think to mention it.

1. I didn't clear it. It automatically shut down and did a restart.
2. I could not finish. This error happened three times in addition to other complications stemming from the arrestee. We went to the hospital.
3. This happened when I chose DUI from the scroll down. Lost all entered data. No sample error, no monkey at the keyboard. You have a real problem and I doubt it can fix itself. I do a lot of DUI arrests. Haven't seen this error before.

Sent from my LG phone

Samuel Beaudreault <sabeaudreault@alaska.edu> wrote:

Officer Brown,

I heard you had a status message of "Fatal Error" while borrowing our Datamaster on 1/28/14.

We haven't had it happen again, so it seems to be a non-issue so far.

I am checking with the crime lab to be sure, but could you answer a few questions?

1. what did you do to clear that status message
2. did you use the same machine to finish the breath test process with that arrestee, or did you go use a different one?
3. were there any other issues you noticed with our machine?

Thanks for the help!

--

Officer Sam Beaudreault
UAF Police Department
ph. [\(907\)474-7721](tel:(907)474-7721)
fax [\(907\)474-5555](tel:(907)474-5555)

--

Officer Sam Beaudreault
UAF Police Department
ph. [\(907\)474-7721](tel:(907)474-7721)
fax [\(907\)474-5555](tel:(907)474-5555)