

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100360

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/04/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.82: 0.080  
LOT #: AG923401  
EXPIRATION: 08/22/2021  
TANK PRESSURE: 1032 psi

BLANK TEST	0.000	16:01
INTERNAL STANDARD	VERIFIED	16:01
EXTERNAL STANDARD	0.077	16:01
BLANK TEST	0.000	16:02
EXTERNAL STANDARD	0.077	16:02
BLANK TEST	0.000	16:03
EXTERNAL STANDARD	0.078	16:03
BLANK TEST	0.000	16:04
EXTERNAL STANDARD	0.078	16:05
BLANK TEST	0.000	16:05
EXTERNAL STANDARD	0.077	16:06

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.4°C	PASSED

### PUMP INFO

Flow Rate = 4.790 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only Jm 06/11/21