

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 6 March 2025

Instrument #: 12-22146

(100365 Jm 06/26/25)

Instrument Location: Eielson AFB

Supervisor Name and Agency: Melvin Hale USAF 354 Security Forces Squadron

Describe the Status Change or Issue with the Instrument:

Replaced Instrument 2-24144 (100424 Jm 06/26/25)

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 6 March 2025 / 2221

Out of Service Date/Time:

For Use by SCDL

Additional Notes

Instrument was never placed in service at the agency due to  
Interference Detected and Internal Standard Error status messages.  
Jm 06/26/25

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor