

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: 05/10/22 Instrument #: 100345

Instrument Location: VALDEZ POLICE DEPT., VALDEZ AK

Supervisor Name and Agency: Chad Clements, UPD

Describe the Status Change or Issue with the Instrument:

TAKING INSTRUMENT OUT OF SERVICE TO SEND TO CRIME LAB FOR MAINTENANCE

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time:

Out of Service Date/Time: 5/10/22 @ 2200

For Use by SCDL

Additional Notes

Instrument pulled from service due to a high prevalence of standard out of range status messages

TLM 08/15/22

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740