

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100349  
-----

Date: 09/25/2019  
Time: 05:44:14

OPERATOR NAME:  
Q Q Q  
CERTIFICATE #: Q

Ca = 0.1000  
CAL = 1.009011    0.800 <= CAL < 1.200  
b1 = 0.0003    0.0000 <= b1 < 0.0040  
b2 = 0.0040    0.0010 <= b2 < 0.0100  
b3 = 0.0003    0.0000 <= b3 < 0.0040  
Xq = 0.0768    0.0500 <= Xq < 0.2000  
a21 = 1.177280    1.050 <= a21 < 1.300  
a31 = 0.557458    0.300 <= a31 < 0.800