

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100364

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/28/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.27: 0.081
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 1194 psi

BLANK TEST	0.000	03:29
INTERNAL STANDARD	VERIFIED	03:29
EXTERNAL STANDARD	0.082	03:29
BLANK TEST	0.000	03:30
EXTERNAL STANDARD	0.082	03:30
BLANK TEST	0.000	03:31
EXTERNAL STANDARD	0.082	03:31
BLANK TEST	0.000	03:32
EXTERNAL STANDARD	0.082	03:32
BLANK TEST	0.000	03:33
EXTERNAL STANDARD	0.082	03:33
BLANK TEST	0.000	03:34

Average = 0.0820
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 40.6°C PASSED

PUMP INFO

Flow Rate = 5.346 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

3/8/18

Subscribed and sworn before me this 08 day of 03, 2018


Carlie K. Bailey, Notary Public
My Commission Expires With Office



