

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100361

Received From: AST-DELTA

Date: 9/3/25

Analyst: Kaja Swisher

Reason for instrument return:

Black screen, internal standard errors. Instrument unable
to turn on, beeping is constant.
(internal short v.a. solder wire)

From the field

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐