

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 7/24/2023

Instrument #: 100392

Instrument Location: Homer Police Department

Supervisor Name and Agency: Trooper David Bozman/AST

Describe the Status Change or Issue with the Instrument:

Instrument was placed in service and a VOC was performed indicating proper operation.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 07/24/2023 1232

Out of Service Date/Time: N/A

For Use by SCDL

Additional Notes

On 07/21/2023, Trooper Bozman advised us that the DMT screen is not functioning. He rebooted the DMT, checked all the external connections, the inside of the DMT and the screen was still not responding.

On 07/24/2023, he reached out again saying the screen is now functioning without him having to do anything to it and it was placed in service. TLM 07/26/2023

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor