

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100370

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/03/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.37: 0.079
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 856 psi

BLANK TEST	0.000	13:42
INTERNAL STANDARD	VERIFIED	13:42
EXTERNAL STANDARD	0.079	13:43
BLANK TEST	0.000	13:44
EXTERNAL STANDARD	0.079	13:44
BLANK TEST	0.000	13:45
EXTERNAL STANDARD	0.079	13:45
BLANK TEST	0.000	13:46
EXTERNAL STANDARD	0.079	13:46
BLANK TEST	0.000	13:47
EXTERNAL STANDARD	0.079	13:47
BLANK TEST	0.000	13:48

Average = 0.0790
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED
Breath Tube = 45.7°C PASSED

PUMP INFO

Flow Rate = 4.686 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Charles R. Foster
Scientific Director
State Breath Alcohol Program

8/10/17

Subscribed and sworn before me this 10 day of Aug, 2017




Kyle Klöpfer, Notary Public
My Commission Expires With Office

COB 8/8/17