

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 08/08/2017
Effective: 08/10/2017

Version: DMEF2017 R1
Status: Active

Instrument # 100665

Date: 12/18/17

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Unable to upload instrument in the field or recognize a USB drive

2. Visual Inspection (internal and external)

3. O-ring inspection

4. Troubleshooting/Repair

Replaced o-ring

Instrument successfully uploaded at UCD2 via modem link

Also tested USB drive by updating software to full version of 5.00 without issue.

5. VOC (attached)

6. Internal Standard value $\leq 3.00\%$ Yes No

7. Verify software version

Return to Factory for Repair

Ready for Calibration

Ready for Certification

Proceed to Close Out

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100665

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/18/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.95: 0.080
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 235 psi

BLANK TEST 0.000 15:00
INTERNAL STANDARD VERIFIED 15:00
Xq = 0.129 (0.21%)
EXTERNAL STANDARD 0.078 15:00
X[1] = 0.0784 (-0.0002) (-0.0034)
BLANK TEST 0.000 15:01
EXTERNAL STANDARD 0.079 15:01
X[1] = 0.0785 (-0.0003) (-0.0026)
BLANK TEST 0.000 15:02
EXTERNAL STANDARD 0.078 15:02
X[1] = 0.0785 (-0.0003) (-0.0032)
BLANK TEST 0.000 15:03
EXTERNAL STANDARD 0.078 15:04
X[1] = 0.0784 (-0.0004) (-0.0024)
BLANK TEST 0.000 15:04
EXTERNAL STANDARD 0.079 15:05
X[1] = 0.0786 (-0.0002) (-0.0025)
BLANK TEST 0.000 15:06

Average = 0.0784
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.5°C PASSED
(44.0°C - 52.0°C)
Breath Tube = 41.8°C PASSED
(38.0°C - 50.0°C)

SETTINGS

Lamp Voltage = 1.70 V PASSED
(0.98 V - 2.30 V)
Cooler Voltage = 1.67 V PASSED
(1.03 V - 2.41 V)
Bias Voltage = 79 V PASSED
(47 V - 111 V)
Chopper Freq = 524 Hz PASSED
(475 Hz - 575 Hz)
Barometer = 29.9 in

PUMP INFO

Flow Rate = 5.233 L/M PASSED
(3.500 L/M - 6.500 L/M)

DETECTOR INFO

PUMP ON PASSED
(0.002 V <= 0.010 V)
PUMP OFF PASSED
(0.002 V <= 0.010 V)

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

Xq = 0.128 (0.62%) (0.00% - 4.00%) PASSED