

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100682

Received From: King Cove

Date: 5/16/24

Analyst: Kaija Swisher

### Reason for instrument return:

"Unable to load ticket database" status messages occurring even after restarting. "Restarted due to fatal error": Unable to start instrument (without getting status messages.)

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload  upload via modem line failed / Removed SD card, retrieved files, uploaded, then replaced with new SD card. Uploaded normally with new SD card.

Update agency information and polling status in DMHost Calling Program

Check for VOCs

### From Intoximeters

Intoximeters Repair Paperwork (attached)