

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 11/10/2021

Version: 2.0

Instrument #: 1003664

Received From: APD

Date: 02-14-22

Analyst: Carlie Blairter / Tatiana Lee

Reason for instrument return:

Failed VOC, pump issues & internal standard errors

Change instrument ~~and case location~~ in JusticeTrax Breath Alcohol Database

returned from local (APD) with no case
Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)