

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 7/16/2021 Instrument Serial Number: 100380

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Jason Cusack, Anchorage Police

Describe the Status Change or Issue with the Instrument:

Instrument 100380 was taken out of service at about 0306 hours this morning on 7/16/2021. I was contacted reference a potentially bad O ring which caused the dry gas standard to leak. I attempted to replace the O ring but that did not have any effect and the dry gas continued to leak. I was unable to figure out what was causing the leak.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 7/16/2021 0306

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740