

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 1/9/2023

Instrument #: 100692

Instrument Location: Anchorage Airport Police

Supervisor Name and Agency: Thomas Hoffman Anchorage Airport Police

Describe the Status Change or Issue with the Instrument:

Printer working and passed a VOC on 1/9/23.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 01/09/2023 @ 0900

Out of Service Date/Time:

-----  
**For Use by SCDL**

Additional Notes

Instrument was taken out of service by agency due to printing issues. Printer has been fixed by the agency and the DataMaster is back in service. TLM 01/12/2023

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor