

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 05/28/2024

Instrument #: 100342

Instrument Location: ASCDL

Supervisor Name and Agency: Kaija Swisher, SCDL working with Christopher Wigner, Bethel PD

Describe the Status Change or Issue with the Instrument:

Removed from service for internal standard errors. Refer to In Field Instrument Review 2024.02.22 - 2024.04.11. KS 5/28/2024.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 09/11/2021

Out of Service Date/Time: 5/20/2024

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor