VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument State of Alaska

Serial #: 100342 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:08/15/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.85: 0.080 LOT #: AG018102 EXPIRATION: 06/29/2022

TANK PRESSURE: 556 psi

BLANK TEST

INTERNAL STANDARD VERIFIED 12:13

0.000 12:13

Gas Flow Error

Diagnostic Check

VERSIONS

DMT: 3.02 PIC: 3.03

Modem: 2.6 Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C

PASSED

Breath Tube = 45.8°C

PASSED

PUMP INFO

Flow Rate = 4.183 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2 Filter 3

PASSED PASSED

INTERNAL STANDARD

PASSED