

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/13/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.22: 0.081
LOT #: AG135004
EXPIRATION: 12/16/2023
TANK PRESSURE: 1246 psi

BLANK TEST	0.000	21:18
INTERNAL STANDARD	VERIFIED	21:18
EXTERNAL STANDARD	0.081	21:18
BLANK TEST	0.000	21:19
EXTERNAL STANDARD	0.083	21:19
BLANK TEST	0.000	21:20
EXTERNAL STANDARD	INTERFERENCE	21:20

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 46.4°C	PASSED

PUMP INFO

Flow Rate = 4.478 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File Only JM 06/30/23