

In Field Instrument Review

Dates: 10/24/2017 – 01/03/2018, Prepared by Colleen O’Bryant

100346 (Fairbanks Airport PD) – 1/8 – Blank Error – no action taken
100349 (Ketchikan PD) – 2/14 – RFI, Pump Error – no action taken
100352 (Anchorage Airport PD) – 8/29 – Pump Error, Interference, Ambient Fail, Diagnostic Fail – Diagnostic fails due to breath tube temperature – working with agency to confirm breath hose seated properly
100355 (Seward PD) – 4/15 – Standard Out of Range – Possibly environmental, has not repeated since November, tank expires 03/25/2018 – no action taken
100360 (AST/Soldotna) – 2/26 – Interference – no action taken
100361 (Unalaska PD) – 5/16 – RFI, Tank Interference, Blank Error – instrument removed from service on 1/16/2018
100363 (Nome PD) – 4/18 – Standard Out of Range, Filter Wheel Error - instrument removed from service on 12/24/2017
100364 (APD) – 7/182 – RFI, Pump Error, Blank Error – no action taken
100368 (AST/Willow) – 1/3 – RFI – no action taken
100375 (Denali NP) – 4/7 – RFI, Ambient Fail, Detector Overflow – all messages occurred during “test/test” samples – no action taken
100378 (AST/Fairbanks) – 4/57 – RFI, Interference, Blank Error – no action taken
100379 (Wasilla PD) – 4/18 – RFI, Pump Error – no action taken
100380 (Kenai PD) – 1/28 – Diagnostic Fail – followed by successful test - no action taken
100384 (Juneau PD) – 1/25 – Standard Out of Range – no action taken
100396 (APD) – 4/99 – RFI, Blank Error, Internal Standard Error – Instrument O/S 12/7/2017
100397 (North Slope Borough PD) – 5/13 – Pump Error, Standard Out of Range – tank change occurred on 12/17/2017 – no further action taken
100406 (Palmer PD) – 3/50 – RFI, Pump Error – no action taken
100413 (Homer PD) – 1/21 – RFI – no action taken
100414 (Eielson AFB) – 6/20 – Gas Flow Error, Ambient Fail – Instrument replaced by agency owned DMT 01/22/2018 – no action taken
100416 (Fairbanks PD) – 4/50 – RFI, Pump error – no action taken
100421 (JBER) – 6/17 – Interference, Ambient Fail, Static Analysis Error – Instrument Removed from service.
100666 (Bethel PD) - 1/14 – Pump Error – no action taken
100667 (Mat/Su West) – 4/31 – RFI – no action taken
100669 (APD) – 8/44 – Interference/Blank Error – monitoring instrument
100674 (Juneau PD) – 2/7 – Standard Out of Range – Tank change on 1/14/2018 – no additional action taken
100689 (Kotzebue PD) – 2/16 – Interference – both subject tests followed by completed tests. No action taken
100694 (Petersburg PD) – 1/3 – Pump Error – no action taken
100695 (Wasilla PD) – 1/14 – Pump Error – no action taken
100696 (Soldotna PD) – 3/16 – Gas Flow Error – Tank change on 12/05/2017 – no further action taken
100700 (Wrangell PD) – 1/15 – Pump Error – no action taken

This review does not include data from AST/Emmonak, AST/Talkeetna and Haines PD

