

	<b>State of Alaska Scientific Crime Detection Laboratory</b>		
	<b>DataMaster Acceptance Form - DMAF2011R0</b>		
	Issuing Authority - Nita Bolz, Scientific Director		
Approval date: 03/01/11	Effective date: 03/01/11	Archive date: Active	

**DataMaster DMT Acceptance Form**

Instrument # 100351

Date: 3/16/11

Analyst: C. O' Bryant

1. Check in instrument in to JusticeTrax Breath Alcohol Database
2. Visual Inspection
3. Update software
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument
7. State of Alaska Property ID# 10140905



## CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dnt: 100351  
-----

Date: 01/11/2011  
Time: 09:23:58

Ca	=	0.1000		
CAL	=	0.996521	0.800	<= CAL < 1.200
b1	=	0.0002	0.0000	<= b1 < 0.0040
b2	=	0.0054	0.0010	<= b2 < 0.0100
b3	=	0.0000	0.0000	<= b3 < 0.0040
Xq	=	0.0975	0.0500	<= Xq < 0.2000
a21	=	1.151480	1.050	<= a21 < 1.300
a31	=	0.476583	0.300	<= a31 < 0.800



*Certificate of Calibration*

100351

Instrument Serial Number

10180

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

1-11-11

Date

A handwritten signature in black ink, appearing to be "C. Bryant", written over a horizontal line.

Technician

Final page for report

\

VANCCLSAN01\JTRAX\CASEDOCS\B-100351\B-100351\_ECF-20140

6/27/2014