

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100672

Analyst(s): Carlie Plaister

Start Date: 1/3/2024

Date Received: 12/1/2023

Reason for instrument evaluation

Preventative Maintenance

Visual Inspection

- | | |
|---|---|
| <input checked="" type="checkbox"/> Clean | <input checked="" type="checkbox"/> Replace missing screws |
| <input checked="" type="checkbox"/> Cermetex Box (modem) is secure | <input checked="" type="checkbox"/> Ensure modem cable present |
| <input checked="" type="checkbox"/> Double Check Valve secured to wall | <input checked="" type="checkbox"/> All tubing and connections secure |
| <input checked="" type="checkbox"/> O-ring inspection and ensure 2 replacement O-rings in bag | |

As Found Conditions:

Barometric Pressure (inHg): 28.98 Temperature (C): 22.6 Relative Humidity (%): 14

Linearity Test (attached)

0.020g/210L Mnf: Airgas	Lot: AG217902	Exp: 10/28/24
0.080g/210L Mnf: Airgas	Lot: AG310901	Exp: 4/19/25
0.300g/210L Mnf: Airgas	Lot: AG327804	Exp: 10/05/25

Notes

Secured double check valve to wall. Replaced o-ring.
1st VOC - blank error & wrong software version. Installed 3.02
Blank error could be from recent use of dust off.

- | | |
|---|--|
| <input checked="" type="checkbox"/> VOC (attached) | <input checked="" type="checkbox"/> Verify software version |
| <input checked="" type="checkbox"/> VOC Internal Standard value ≤ 3.00% | <input checked="" type="checkbox"/> Non-Drinking Subject Test (attached) |

Condition of item: Good Needs Repair

- | |
|---|
| <input type="checkbox"/> Ready for Adjustment |
| <input checked="" type="checkbox"/> Ready for Calibration |

End Date: 1/3/2024

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

LINEARITY CHECK

0.600

Alaska Department of Public Safety
DATAMASTER dmt: 100672

Date: 01/03/2024
Time: 13:04:56

SUPERVISOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436

TARGET	AVERAGE	STD.	DEV.
0.019	0.0186	0.0005	
0.077	0.0758	0.0004	
0.290	0.2854	0.0005	

SLOPE = 0.984376
 R^2 = 1.000000

0.300

0.200

0.100

0.000

0.600 0.500 0.400 0.300 0.200 0.100 0.000

CW

LINERARITY CHECK: Control Test 1

Alaska Department of Public Safety
DATAMASTER dmt: 100672






Date: 01/03/2024
Time: 13:04:57

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020
TARGET AT 28.97: 0.019
LOT #: AG217902
EXPIRATION: 06/28/2024
TANK PRESSURE: 554 psi

BLANK TEST	0.000	13:06
INTERNAL STANDARD	VERIFIED	13:06
EXTERNAL STANDARD	0.018	13:06
BLANK TEST	0.000	13:07
EXTERNAL STANDARD	0.019	13:07
BLANK TEST	0.000	13:08
EXTERNAL STANDARD	0.019	13:08
BLANK TEST	0.000	13:09
EXTERNAL STANDARD	0.019	13:09
BLANK TEST	0.000	13:10
EXTERNAL STANDARD	0.018	13:10
BLANK TEST	0.000	13:11

Average = 0.0186
Std Dev = 0.0005



LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety
DATAMASTER dmt: 100672

Date: 01/03/2024
Time: 13:12:39


EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.96: 0.077
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 463 psi

BLANK TEST	0.000	13:13
INTERNAL STANDARD	VERIFIED	13:13
EXTERNAL STANDARD	0.075	13:14
BLANK TEST	0.000	13:15
EXTERNAL STANDARD	0.076	13:15
BLANK TEST	0.000	13:16
EXTERNAL STANDARD	0.076	13:16
BLANK TEST	0.000	13:17
EXTERNAL STANDARD	0.076	13:17
BLANK TEST	0.000	13:18
EXTERNAL STANDARD	0.076	13:18
BLANK TEST	0.000	13:19

Average = 0.0758

Std Dev = 0.0004



LINEARITY CHECK: Control Test 3

Alaska Department of Public Safety
DATAMASTER dmt: 100672






Date: 01/03/2024
Time: 13:21:46

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.300
TARGET AT 28.94: 0.290
LOT #: AG327804
EXPIRATION: 10/05/2025
TANK PRESSURE: 352 psi

BLANK TEST	0.000	13:22
INTERNAL STANDARD	VERIFIED	13:23
EXTERNAL STANDARD	0.285	13:23
BLANK TEST	0.000	13:24
EXTERNAL STANDARD	0.285	13:24
BLANK TEST	0.000	13:25
EXTERNAL STANDARD	0.286	13:25
BLANK TEST	0.000	13:26
EXTERNAL STANDARD	0.285	13:26
BLANK TEST	0.000	13:27
EXTERNAL STANDARD	0.286	13:27
BLANK TEST	0.000	13:28

Average = 0.2854
Std Dev = 0.0005



VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100672

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/03/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.99: 0.077
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 481 psi

BLANK TEST	0.000	12:44
INTERNAL STANDARD	VERIFIED	12:44
EXTERNAL STANDARD	0.074	12:45

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.01
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C
Breath Tube = 46.0°C

PASSED
PASSED

PUMP INFO

Flow Rate = 4.817 L/M

PASSED

DETECTOR INFO

PUMP ON
PUMP OFF

PASSED
PASSED

FILTER INFO

Filter 1
Filter 2
Filter 3

PASSED
PASSED
PASSED

INTERNAL STANDARD

PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this _____ day of _____, 20 _____

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100672

Date: 01/03/2024
Time: 13:32:07

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436

DEPT/AGENCY:
NON-DRINKING SUBJECT TEST
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.95: 0.077
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK	PASSED	13:33
BLANK TEST	0.000	13:33
INTERNAL STANDARD	VERIFIED	13:33
EXTERNAL STANDARD	0.076	13:33
BLANK TEST	0.000	13:34
SUBJECT SAMPLE (Vol=1.94L)	0.000	13:34
BLANK TEST	0.000	13:35
EXTERNAL STANDARD	0.076	13:36
BLANK TEST	0.000	13:36

40.00

36.00

32.00

28.00

24.00

20.00

16.00

12.00

8.00

4.00

0.00

25

10

5

0