

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100668

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/09/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.74: 0.080
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 548 psi

BLANK TEST	0.000	03:36
INTERNAL STANDARD	VERIFIED	03:36
EXTERNAL STANDARD	0.079	03:36
BLANK TEST	0.000	03:37
EXTERNAL STANDARD	0.079	03:37
BLANK TEST	0.000	03:38
EXTERNAL STANDARD	0.079	03:39
BLANK TEST	0.000	03:39
EXTERNAL STANDARD	0.079	03:40
BLANK TEST	0.000	03:41
EXTERNAL STANDARD	0.079	03:41
BLANK TEST	0.000	03:42

Average = 0.0790
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 4.862 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only DMV 03/28/23