

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100350

Date: 3/29/16

Analyst: CWF

1. Reason for instrument evaluation

Used for training at APD

2. Visual Inspection

3. Troubleshooting/Repair

None

4. VOC (attached)

5. Internal Standard value $\leq 3.00\%$ Yes No

6. Verify software version

Return to Factory for Repair

Ready For Calibration

Ready For Certification

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100350

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/30/2016

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 29.92: 0.080
 LOT #: 09214080A2
 EXPIRATION: 05/01/2016
 TANK PRESSURE: 490 psi

BLANK TEST	0.000	09:14
INTERNAL STANDARD	VERIFIED	09:14
Xq = 0.098 (1.92%)		
EXTERNAL STANDARD	0.078	09:14
X[1] = 0.0782 (-0.0009) (-0.0002)		
BLANK TEST	0.000	09:15
EXTERNAL STANDARD	0.078	09:15
X[1] = 0.0777 (-0.0002) (0.0011)		
BLANK TEST	0.000	09:16
EXTERNAL STANDARD	0.078	09:17
X[1] = 0.0775 (0.0004) (-0.0002)		
BLANK TEST	0.000	09:17
EXTERNAL STANDARD	0.078	09:18
X[1] = 0.0778 (0.0003) (-0.0006)		
BLANK TEST	0.000	09:19
EXTERNAL STANDARD	0.078	09:19
X[1] = 0.0779 (0.0005) (-0.0006)		
BLANK TEST	0.000	09:20

Average = 0.0780
 Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
 PIC: 3.02
 Modem: 2.6
 Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 47.2°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.78 V	PASSED
(1.10 V - 2.56 V)	
Cooler Voltage = 1.82 V	PASSED
(1.01 V - 2.37 V)	
Bias Voltage = 79 V	PASSED
(47 V - 111 V)	
Chopper Freq = 524 Hz	PASSED
(475 Hz - 575 Hz)	

PUMP INFO

Flow Rate = 4.410 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

Xq = 0.099	1.66%	(0.00% - 4.00%)
------------	-------	-----------------