

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 2-19-20 Instrument Serial Number: 100416

Instrument Location: Fairbanks PD

Supervisor Name and Agency: Lane Bonham Fairbanks PD

Describe the Status Change or Issue with the Instrument:

When running a subject test, the Instrument froze and tried to restart on its own. The Instrument again froze and had to be turned off and then back on.
Once the Instrument warmed up, it was ready for a subject test. Once the subject test was started, everything began to work correctly, until it
attempted to run the External Standard, the Instrument stopped and stated it had Interference Detected Status Message on two different subject samples.
A VCO was attempted and on the External Standard test, it had another Interference Detected Status Message. The tank was removed and then
replaced, which fixed the issue. A successful VCO and Subject Test was completed.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 2-8-20 0422 hours

Out of Service Date/Time: 2-8-20 0533 hours

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740