

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100382

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/15/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.59: 0.079
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 870 psi

BLANK TEST	0.000	15:50
INTERNAL STANDARD	VERIFIED	15:50
EXTERNAL STANDARD	0.080	15:50
BLANK TEST	0.000	15:51
EXTERNAL STANDARD	0.079	15:51
BLANK TEST	0.000	15:52
EXTERNAL STANDARD	0.080	15:52
BLANK TEST	0.000	15:53
EXTERNAL STANDARD	0.080	15:53
BLANK TEST	0.000	15:54
EXTERNAL STANDARD	0.080	15:54
BLANK TEST	0.000	15:55

Average = 0.0798
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 47.7°C	PASSED

PUMP INFO

Flow Rate = 5.065 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File only. Dhw 12/15/20