

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100348

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 04/30/2024

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.07: 0.080  
LOT #: AG310901  
EXPIRATION: 04/19/2025  
TANK PRESSURE: 999 psi

BLANK TEST	0.000	07:21
INTERNAL STANDARD	VERIFIED	07:21
EXTERNAL STANDARD	0.075	07:21
BLANK TEST	0.000	07:22
EXTERNAL STANDARD	0.075	07:23
BLANK TEST	0.000	07:23
EXTERNAL STANDARD	0.075	07:24
BLANK TEST	0.000	07:25
EXTERNAL STANDARD	0.075	07:25
BLANK TEST	0.000	07:26
EXTERNAL STANDARD	0.075	07:26
BLANK TEST	0.000	07:27

Average = 0.0750  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 44.8°C	PASSED

### PUMP INFO

Flow Rate = 4.434 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only JMW 05/07/24