

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 12/27/2020 Instrument Serial Number: 100385

Instrument Location: Heenah Alaska

Supervisor Name and Agency: Joshua Huskey Heenah PD

Describe the Status Change or Issue with the Instrument:
Faulty File transfer. Was recommended to switch instruments

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 12/27/2020 1042 hrs

For Use by SCDL

Additional Notes

Some .zip files were not able to be located after output to USB. Instrument returned to Crime Lab to reconcile records.
JHR 01/11/2021

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740