

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100395

Received From: AST/Aniak

Date: 5/19/25

Analyst: Carlie Glaister

Reason for instrument return:

Filter wheel errors & communication error.

### From the field

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

### From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐