

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100383

Date: 2/1/16

Analyst: CRF

1. Reason for instrument evaluation

Communication Error / SCTemp: 0.0°C message

2. Visual Inspection

3. Troubleshooting/Repair

Replaced ribbon cable connecting controller board to
SI Board - Didn't solve problem - put original back

Replaced controller board - Didn't solve problem - put
original back

4. VOC (attached) Can't initiate - SCTemp: 0.0°C

5. Internal Standard value $\leq 3.00\%$ Yes No N/A

6. Verify software version - Can't access menu options

Return to Factory for Repair

Ready For Calibration

Ready For Certification

} N/A

Disassembling for
Supervisor Class
demonstration - CRF
8/4/16